



## ISTITUTO NAZIONALE DI RICERCA METROLOGICA Repository Istituzionale

Corrigendum: Refractive index gas thermometry between 13.8 K and 161.4 K (2021 Metrologia

*Original*

Corrigendum: Refractive index gas thermometry between 13.8 K and 161.4 K (2021 Metrologia

58 025008) / Madonna Ripa, D; Imbraguglio, D; Gaiser, C; Steur, P P M; Giraudi, D; Fogliati, M; Bertinetti, M; Lopardo, G; Dematteis, R; Gavioso, R M. - In: METROLOGIA. - ISSN 0026-1394. - 58:6(2021), p. 069501. [10.1088/1681-7575/ac2d9e]

*Availability:*

This version is available at: 11696/73212 since: 2022-02-18T15:01:37Z

*Publisher:*

IOP Publishing Ltd

*Published*

DOI:10.1088/1681-7575/ac2d9e

*Terms of use:*

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

*Publisher copyright*

(Article begins on next page)



---

**CORRIGENDUM • OPEN ACCESS**

## Corrigendum: Refractive index gas thermometry between 13.8 K and 161.4 K (2021 *Metrologia* [58 025008](#))

To cite this article: D Madonna Ripa *et al* 2021 *Metrologia* **58** 069501

View the [article online](#) for updates and enhancements.

## Corrigendum

# Corrigendum: Refractive index gas thermometry between 13.8 K and 161.4 K (2021 Metrologia 58 025008)

D Madonna Ripa<sup>1</sup>, D Imbraguglio<sup>1</sup>, C Gaiser<sup>2</sup>, P P M Steur<sup>1</sup>,  
D Giraudi<sup>1</sup>, M Fogliati<sup>1</sup>, M Bertinetti<sup>1</sup>, G Lopardo<sup>1</sup>, R Dematteis<sup>1</sup> and  
R M Gavioso<sup>1,\*</sup>

<sup>1</sup> Istituto Nazionale di Ricerca Metrologica, Strada delle Cacce 91, 10135 Torino, Italy

<sup>2</sup> Physikalisch-Technische Bundesanstalt (PTB), Abbestrasse 2-12, 10587 Berlin, Germany

E-mail: [r.gavioso@inrim.it](mailto:r.gavioso@inrim.it)

Received 1 September 2021, revised 23 September 2021

Accepted for publication 7 October 2021

Published 25 October 2021



CrossMark

A mass calibration of the weights set of the pressure balance used to provide the primary pressure reference to the refractive index gas thermometer (RIGT) revealed some relevant discrepancies from the mass calibration data made available from the manufacturer. Particularly, the 0.5 kg mass carrying bell in the weights set was found to be lighter than previously assumed by  $(45 \pm 3)$  mg.

As a consequence, the experimental pressure records of the RIGT measurements must be corrected by variable amounts ranging between  $-1.5$  Pa and  $-2.4$  Pa depending on the particular combination of weights employed to realize pressures in the range between 40 kPa and 380 kPa.

Upon applying these corrections, thermodynamic temperatures have been re-evaluated leading to the revised  $(T - T_{90})$  estimates listed below in a corrected version of table 1:


**Table 1.** Recommended  $(T - T_{90})$  differences based on results obtained using two thermometric gases.

$T_{90}/\text{K}$	He	Ne	Recommended
	$(T - T_{90})/\text{mK}$		
13.8033	$0.75 \pm 1.70$		<b><math>0.75 \pm 1.70</math></b>
24.5561	$-1.11 \pm 0.39$		<b><math>-1.11 \pm 0.39</math></b>
54.3584	$-3.16 \pm 0.53$	$-4.94 \pm 1.23$	<b><math>-3.44 \pm 0.53</math></b>
83.8058	$-3.61 \pm 1.05$	$-5.16 \pm 1.10$	<b><math>-4.35 \pm 1.05</math></b>
161.4060	$-5.30 \pm 2.90$	$-8.77 \pm 4.45$	<b><math>-6.37 \pm 2.90</math></b>

## Acknowledgments





This work was carried out within the projects ‘Implementing the new Kelvin 2—InK2’ and ‘Realising the New Kelvin—Real K’. These projects (15SIB02 InK 2) and (18SIB02 Real K) have received funding from the EMPIR programme which is co-financed by the Participating States and from the European Union’s Horizon 2020 research and innovation programme.

\* Author to whom any correspondence should be addressed.

 Original content from this work may be used under the terms of the [Creative Commons Attribution 4.0 licence](https://creativecommons.org/licenses/by/4.0/). Any further distribution of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI.

**ORCID iDs**

D Madonna Ripa  <https://orcid.org/0000-0002-8760-8574>  
D Imbraglio  <https://orcid.org/0000-0002-1289-772x>

C Gaiser  <https://orcid.org/0000-0003-1745-7368>  
P P M Steur  <https://orcid.org/0000-0001-5161-8554>  
G Lopardo  <https://orcid.org/0000-0002-5250-0459>  
R M Gavioso  <https://orcid.org/0000-0002-1631-5133>