



ISTITUTO NAZIONALE DI RICERCA METROLOGICA Repository Istituzionale

NRIM-BIPM.EM-K13.a and b Key Comparisons of Electrical Resistance Standards using 1 Ω and 10 k Ω Resistors as Travelling Standards: INRIM results

Original

NRIM-BIPM.EM-K13.a and b Key Comparisons of Electrical Resistance Standards using 1 Ω and 10 k Ω Resistors as Travelling Standards: INRIM results / Capra, PIER PAOLO; Galliana, Flavio; Francese, Claudio; Lanzillotti, Marco; RONCAGLIONE TET, Luca. - (2023), pp. 1-12.

Availability:

This version is available at: 11696/75319 since: 2023-01-27T09:57:01Z

Publisher:

Published

DOI:

Terms of use:

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

(Article begins on next page)

*Pier Paolo Capra, Flavio Galliana, Claudio Francese,
Marco Lanzillotti, Luca Roncaglione Tet*

**INRIM-BIPM.EM-K13.a and b Key Comparisons of Electrical Resistance
Standards using 1 Ω and 10 k Ω Resistors as Travelling Standards:
INRIM results**

T.R. 3/2023

January 2023

,67,7872 1\$=,21\$/(', 5,&(5&\$ 0(752/2*,8
5HSRVLWRU\ ,VWLWX]LRQDOH

15,0 %,30 (0 . D DQG E .H\ &RPSDULVRQV RI (OHFWULFDO 5HVLV
DQG N¼ 5HVLVWRUV DV 7UDYHOOLQJ 6WDQGDUGV ,15,0 UHVXOWV

2ULJLQDO
15,0 %,30 (0 . D DQG E .H\ &RPSDULVRQV RI (OHFWULFDO 5HVLVWDQFH 6W
5HVLVWRUV DV 7UDYHOOLQJ 6WDQGDUGV ,15,0 UHVXOWV &DSUD 3,(5 3\$
/DQ]LOORWWL 0DUFR 521&\$*/,21(7(7 /XFD SS

\$YDLODELOLW\
7KLV YHUVLRQ LV DYDLODEOH DW VLQFH 7 =

3XEOLVKHU

3XEOLVKHG
'2,

7HUPV RI XVH

7KLV DUWLFOH LV PDGH DYDLODEOH XQGHU WHUPV DQG FRQGLWLRQV DV V
GHVFULSWLRQ LQ WKH UHSRVLWRU\

3XEOLVKHU FRS\ULJKW

\$UWLFOH EHJLQV RQ QH[W SDJH

Table 1. Travelling resistors to be calibrated in the comparison.

<i>Standard</i>	<i>Manufacturer</i>	<i>Mod.</i>	<i>no.</i>
STD 1 Ω	CSIRO	NML	64200
STD 1 Ω	CSIRO	NML	64203
STD 10 k Ω	TEGAM	SR104 - B10K09	K205039730104
STD 10 k Ω	TEGAM	SR104 - B10K12	K201089830104

The standards were carefully packed, but when the container arrived at INRIM one of the sides had been deformed during the travel (Fig. 1 a, and b). The condition of the container was promptly reported to the BIPM.



Fig. 1a, b – Unboxing and check

3. Climatic conditioning of the standards – oil and air baths

The 1 Ω resistors were placed in a Guildline 9732 VT no. 58.805 oil bath with a set point temperature of 23,000°C (Fig. 2). The oil temperature is measured with a calibrated Leeds & Northrup 25 Ω s/ 177 3055 thermometer. The depth of the resistors in the oil, measured between the surface and the top of the standards was 32 mm as indicated in Fig. 3.



Fig. 2 - Guildline 9732 VT oil bath. The short-term stability of the bath is approximately 0.001°C

The 10 k Ω resistors were housed inside a Kambic air bath mod. TK-105 US no. 15115002 set at the temperature of 23°C. The air temperature near the two resistors was measured with a calibrated 100 Ω no. 4446 FLUKE thermometer.




Fig. 3 - Oil immersion depth of the 1 Ω standards.




Fig. 4 - 10 k Ω standards air-bath. The resistors were placed inside a Kambic air-bath mod. 105 with a set-point of 23.000°C and a short-term stability of approximately 0.005°C. The temperature was monitored with a PT100 platinum thermometer.

All the resistors were connected to the measuring instrumentation by means of Teflon-insulated twisted AXON-type cables equipped with copper termination forks. The temperatures of the two baths were measured by always carrying out two series of resistance measurements in succession and at two different current values (1 mA and 1.41 mA) and making the correction due to self-heating. In correspondence with the resistance measurements of the traveling standards, the atmospheric pressure was also acquired using a Barocap-Technology instrument.

4. Main instrumentation – Measurements system

The measurement system used by INRIM for this international comparison is described in detail in the INRIM technical procedure [2]. Nevertheless, the same system is still based on the one described in [3] with only little difference as the use of a new high-performance

automated CC bridge in place of the previous manually operated bridge. The group of INRIM working standards consists of the resistors in fig. 5. The arrows indicated in the diagram show the comparisons and the direction of the traceability in the measurements for the calibration of all the INRIM standards and for the periodic checks. The origin of traceability is the QHE that is used for the calibration of a material transfer resistor, used for the calibration of a 12960 Ω /10 Ω Hamon network. Subsequently, by means of ratios 1:10 directly with a Measurements International MI 6010Q (DCC) bridge, or in ratio 1:100 with Hamon and DCC networks, all the elements of the scale are calibrated. The first step of the precision measurements was carried out during week No. 46 with a calibration of the QHE-10 Ω transfer Hamon network with an associated uncertainty equal to $U = 7 \times 10^{-7}$ at 2σ confidence level. The transfer of traceability from the 10 Ω Hamon network to the first element of the decade scale must take place within a short time, usually a few hours. The first group of resistors to receive traceability consists of the 10 Ω – 1 Ω – 100 Ω standards, which are calibrated and compared with each other both in a 1:10 ratio and with the 10 Ω Hamon network. Due to the impossibility of exchanging the compared resistors when they are in a ratio of 1:10, to reduce the bridge interchange error, the Hamon network is used by comparing the resistors following the diagram in Fig. 5, which allows reducing this type of systematic error. The second group of resistors consists of 100 Ω – 1k Ω – 10k Ω . Traceability is transferred to the 10 k Ω resistor via a 1k Ω step Hamon network, while the 1 k Ω resistor is calibrated against both 100 Ω and 10 k Ω , again in a 1:10 ratio. The precision measurements on the K13 traveling standards are carried out in a 1:1 ratio with 1 Ω and 10 k Ω standards of the INRIM working standard scale.

Annex

In table 5 the measurement results provided by INRIM to the BIPM for the comparison are listed along with the corresponding dates, temperature and pressure values.

Table. 5 –INRIM measurement results in Excel format along with the corresponding dates, temperature and pressure values.

INRIM
EM.K13 a & b
2022

	07-nov	08-nov	09-nov	10-nov	11-nov	12-nov	13-nov	14-nov	15-nov	16-nov	17-nov	23-nov	24-nov	25-nov	26-nov	27-nov
NML 64200 - 50mA	0.999999222	0.999999231	0.999999235	0.999999249	0.999999248	0.999999246	0.999999251	0.999999176	0.999999255	0.999999262	0.999999254	0.999999201	0.999999181	0.999999167	0.999999208	0.999999197
NML 64200 - 100 mA	0.999999204	0.999999212	0.999999218	0.999999236	0.999999236	0.999999229	0.999999237	0.999999232	0.999999235	0.999999245	0.999999237	0.999999206	0.999999177	0.999999182	0.999999187	0.999999176
NML 64203 - 50 mA	1.000000466	1.000000477	1.000000484	1.000000505	1.000000509	1.000000505	1.000000511	1.000000511	1.000000511	1.000000532	1.000000516	1.000000482	1.00000046	1.000000453	1.000000419	1.000000445
NML 64203 - 100 mA	1.00000046	1.000000466	1.000000467	1.000000485	1.000000497	1.000000483	1.000000498	1.000000494	1.000000496	1.000000503	1.000000494	1.000000459	1.000000438	1.000000442	1.000000432	1.000000422
T oil bath (°C)	23.001	23.001	23.001	23.000	23.001	23.001	23.002	23.001	23.002	23.001	23.001	23.001	23.001	22.999	22.998	22.999
P (hPa)	991.4	988.0	990.0	992.0	1005.0	1003.5	995.5	993.7	989.9	978.1	974.0	972.8	983.7	989.5	995.4	1000.0
	07-nov	08-nov	09-nov	10-nov	11-nov	12-nov	13-nov	14-nov	15-nov	16-nov	17-nov	23-nov	24-nov	25-nov	26-nov	27-nov
B10K09	10000.00052	10000.00025	10000.00041	10000.00048	10000.00051	10000.00021	10000.00022	9999.99984	10000.00019	10000.00017	10000.00011	9999.999979	10000.00014	10000.0003	10000.00013	10000.00005
B10K12	10000.01144	10000.01097	10000.01093	10000.01121	10000.01118	10000.01103	10000.01092	10000.01067	10000.01095	10000.01081	10000.01068	10000.01102	10000.01097	10000.01097	10000.01091	10000.01084
T air bath (°C)	22.959	22.960	22.980	22.980	22.981	22.980	22.981	22.980	22.980	22.979	22.979	22.976	22.980	22.979	22.979	22.980
P (hPa)	991.91	990.60	990.50	994.00	1005.81	1003.42	998.37	995.31	989.52	977.52	974.93	975.06	985.56	991.23	997.57	1000.08

INRIM
EM.K13 a & b
2022

NML 64200 - 50mA
NML 64200 - 100 mA
NML 64203 - 50 mA
NML 64203 - 100 mA
T oil bath (°C)
P (hPa)

B10K09
B10K12
T air bath (°C)
P (hPa)

	28-nov	29-nov	30-nov	01-dic	02-dic	03-dic	04-dic	05-dic	07-dic	08-dic	09-dic	10-dic	11-dic	12-dic
NML 64200 - 50mA	0.999999203	0.999999224	0.999999228	0.999999213	0.999999211	0.999999193	0.999999241	0.999999243	0.999999224	0.999999172	0.999999215	0.999999224	0.999999227	0.999999206
NML 64200 - 100 mA	0.999999184	0.99999921	0.999999206	0.999999205	0.9999992	0.999999207	0.999999216	0.99999923	0.9999992	0.999999194	0.999999201	0.99999921	0.999999203	0.999999187
NML 64203 - 50 mA	1.00000046	1.000000469	1.000000486	1.000000459	1.000000465	1.000000479	1.000000503	1.000000497	1.00000047	1.000000456	1.00000048	1.000000479	1.000000471	1.000000444
NML 64203 - 100 mA	1.000000432	1.000000454	1.000000456	1.000000446	1.000000452	1.000000463	1.000000464	1.000000477	1.000000447	1.000000436	1.00000045	1.000000453	1.000000456	1.000000426
T oil bath (°C)	22.998	22.995	22.996	22.996	22.997	22.996	22.998	22.998	22.999	22.998	22.998	22.998	22.998	22.999
P (hPa)	991.5	997.3	989.0	989.9	989.3	989.0	984.5	988.3	986.5	982.0	976.3	970.5	970.2	978.0
	28-nov	29-nov	30-nov	01-dic	02-dic	03-dic	04-dic	05-dic	07-dic	08-dic	09-dic	10-dic	11-dic	12-dic
B10K09	9999.999812	10000.00012	10000.00013	9999.999836	10000.00021	10000.00002	10000.00019	9999.999904	9999.999989	10000.0001	9999.999754	9999.99974	10000.00007	10000.00027
B10K12	10000.01082	10000.0111	10000.01074	10000.01094	10000.01088	10000.01102	10000.01089	10000.01071	10000.01076	10000.01106	10000.01084	10000.01084	10000.01107	10000.01114
T air bath (°C)	22.979	22.980	22.979	22.979	22.979	22.979	22.980	22.980	22.979	22.979	22.980	22.979	22.979	22.978
P (hPa)	990.87	998.17	989.50	990.59	989.94	989.15	984.44	989.34	987.00	982.00	975.50	970.50	972.00	980.00